

## Detection of anti-*Trichinella* antibodies in human serum by indirect ELISA

### INDEX

1. Aim and field of application	2
2. Principle of the method	2
3. References	2
4. Definitions	2
5. Devices/instruments	2
6. Reagents and chemicals	3
7. Procedure	4
7.1 <i>Test and control samples preparation</i>	4
7.2 <i>Procedure</i>	4
8. Interpretation of the results	5
9. Characteristics of the method	5
10. Safety measures	5

ANNEX 1 Production of excretory/secretory antigens from muscle larvae of *Trichinella spiralis*

## 1. Aim and field of application

To determine the presence of anti-*Trichinella* sp. antibodies by an enzyme linked immunosorbent assay (ELISA) in human sera.

The method can be used for the serological diagnosis of human trichinellosis.

## 2. Principle of the method

A 96-well microtiter polystyrene plate is coated with *Trichinella spiralis* excretory/secretory (E/S) antigens partially purified in conditions capable of maintaining the antigens' native form.

Control and test serum samples, properly diluted, are distributed in the wells, allowing any anti-*Trichinella* sp. antibodies that are present to bind to the adsorbed antigen.

The antibodies that do not bind to the antigen are eliminated by washing; peroxidase conjugated anti-human IgG goat antibody is then added to each well. This second incubation allows the conjugate to bind to the human antibodies that were bound to the antigens onto the well surface.

The excess conjugate is eliminated by washing, and the activity of the enzyme bound to the human antibodies is measured by adding a chromogen substrate. After incubation, the intensity of the developed color is determined by a spectrophotometer.

The result is interpreted comparing the color intensity of the wells containing the test sera with those containing the controls.

## 3. References

Gómez-Morales MA, Ludovisi A, Amati M, Cherchi S, Pezzotti P, Pozio E. 2008. Validation of an enzyme-linked immunosorbent assay for diagnosis of human trichinellosis. *Clin Vaccine Immunol.* 5:1723-9

OIE/World Organisation for Animal Health, 2018. Trichinellosis (infection with *Trichinella* spp.), Chapter 3.1.20. Manual of Diagnostic Tests and Vaccines for Terrestrial Animals. World Organization for Animal Health [www.oie.int/eng/normes/mmanual/A\\_00013.htm](http://www.oie.int/eng/normes/mmanual/A_00013.htm).

Bruschi F, Gómez-Morales MA, Hill D. 2019. International Commission on Trichinellosis: Recommendations on the use of serological tests for the detection of *Trichinella* infection in animals and humans. *Food and Waterborne Parasitology*, <https://doi.org/10.1016/j.fawpar.2018.e00032>. DM March 3, 2005 (Italian Official Journal. April 13, 2005 n. 85).

## 4. Definitions

ELISA	Enzyme Linked Immunosorbent Assay
Ag	Antigen
Ab	Antibodies
Ag E/S	Excretory/Secretory antigens
BSA	Bovine Serum Albumin

## 5. Devices/instruments

The following instruments are needed to prepare the reagents to perform the ELISA procedure.

Adjustable pipettes (volumes: 1÷1000 µL)

Balance (0.01-100gr)

Automatic plate washer (strongly recommended)

ELISA plate microtiter reader, 450 nm

Freezer ≤ -15°C

Freezer  $\leq -50^{\circ}\text{C}$   
Ice maker  
Incubator  $37 \pm 1^{\circ}\text{C}$   
Magnetic stirrer  
Adjustable volume dispenser (e.g. Multipette Eppendorf®)  
pH meter  
Pipette aid  
Refrigerator  $1\div 8^{\circ}\text{C}$   
Vortex

## 6. Reagents and chemicals

The step-by-step procedure for preparing the reagents is described below.

- 6.1 Analytical grade water
- 6.2 Phosphate buffered saline (PBS), pH  $7.3 \pm 0.2$
- |                           |               |
|---------------------------|---------------|
| $\text{KH}_2\text{PO}_4$  | 0.34 g        |
| $\text{Na}_2\text{HPO}_4$ | 1.21 g        |
| NaCl                      | 8.0 g         |
| Analytical grade water    | up to 1000 mL |

Dissolve compounds in 750 mL of analytical grade water under magnetic stirring. Check the pH ( $7.3 \pm 0.2$ ) and then bring the solution to the final volume; refrigerate. The solution must be used within 6 months.

- 6.3 Carbonate buffered saline, pH  $9.6 \pm 0.2$
- |                          |               |
|--------------------------|---------------|
| $\text{Na}_2\text{CO}_3$ | 1.12g         |
| $\text{NaHCO}_3$         | 2.92g         |
| Analytical grade water   | up to 1000 mL |

Dissolve the compounds in 750 mL of analytical grade water under magnetic stirring. Check the pH ( $9.6 \pm 0.2$ ) and then bring the solution to the final volume; store at room temperature. If needed, clear the solution by filtration. The solution must be used within 6 months.

- 6.4 Washing solution
- |                        |               |
|------------------------|---------------|
| Tween 20               | 1 mL          |
| Analytical grade water | up to 2000 mL |

The solution shall be prepared immediately before use, as follows: add 1 mL of Tween 20 to a 2 L flask; bring the solution to the final volume by adding analytical grade water and mix by magnetic stirring until the solution is clear. If refrigerated, the solution should be used within 24 h.

- 6.5 Blocking solution
- |          |                |
|----------|----------------|
| BSA      | 0.25 g         |
| Tween 20 | 0.05 mL        |
| PBS      | up to 50.00 mL |

The solution shall be prepared immediately before use, as follows: place 0.25 g BSA (bovine serum albumin) directly in a 50 mL tube; add 40 mL of PBS buffer and mix by vortexing until the BSA is completely dissolved. Add 0.05 mL Tween 20; mix by vortexing and bring to volume. If refrigerated, the solution must be used within 24 h.

- 6.6 Sera and conjugate diluter
- |     |       |
|-----|-------|
| BSA | 1.00g |
|-----|-------|



- 7.2.6 Add 100 $\mu$ L of each diluted sample per well and incubate for 30 min at 37°C.
- 7.2.7 Each serum dilution shall be performed in duplicate.
- 7.2.8 Sera shall be diluted 1/200.
- 7.2.9 Wash 3 times in the automatic plate washer with the washing solution.
- 7.2.10 Add 100 $\mu$ L of the diluted anti-human IgG peroxidase labelled antibodies per well and incubate for 1h at 37°C.
- 7.2.11 Wash 3 times in the automatic plate washer with the washing solution.
- 7.2.12 Add 100 $\mu$ L TMB substrate per well; incubate for 10 min at room temperature.
- 7.2.13 Stop the reaction by adding 50 $\mu$ L of the stop solution per well and read the reaction in the ELISA plate microtiter reader at 450nm.

## 8. Interpretation of the results

- 8.1 The test results can be considered as valid if all of the following criteria are fulfilled:
  - 8.1.1 The OD (Optical Density) value of the negative control sera should be lower than the *cut off* value determined by the validation method (i.e., 0.230);
  - 8.1.2 The OD value of the positive control sera has to be >1.0 of the unit absorbance;
  - 8.1.3 The difference in OD between the 2 measures made on the same positive control sample in strict conditions of repeatability has to be  $\leq 0.15$  unit absorbance, and on the same negative control sample it has to be  $\leq 0.05$  unit absorbance.

If even only one of the above-reported criteria is not met, the test has to be considered as non-valid and the sera should be tested again.

- 8.2 Calculate the mean of the 2 duplicates for each positive sera (PS) and for each test sera (TS).
- 8.3 Subtract from each mean value the mean OD value of the blanks (OD<sub>b</sub>).
- 8.4 Select the higher OD value among the positive control sera (PS<sub>max</sub>), and for each sample calculate the extinction value (I<sub>e</sub>) according to the following formula:

$$I_e (\%) = \frac{\text{OD mean duplicates TS} - \text{OD}_b}{\text{OD mean duplicates highest PS} - \text{OD}_b} \times 100\%$$

where:  $I_e \geq 11.8\%$ , *Trichinella* positive serum  
 $I_e < 11.8\%$ , *Trichinella* negative serum

## 9. Characteristics of the method

This method was validated in terms of diagnostic sensitivity, specificity and repeatability. The results of the validation procedure confirmed that the method is suitable for the expected aim and field of application.

## 10. Safety measures

This method has to be carried out only by authorized personnel. The operator should wear individual protection devices during the test performance. For the general safety measures, refer to the guidelines of CDC.

## ANNEX 1

### Production of excretory/secretory antigens from muscle larvae of *Trichinella spiralis*

#### 1 Aim and field of application

To describe the production of excretory/secretory antigens from muscle larvae of *Trichinella spiralis*.

The product can be used as antigens for serology to detect anti-*Trichinella* spp. antibodies.

#### 2 References

Gamble HR, Anderson WR, Graham CE, Murrell KD. 1983. Diagnosis of swine trichinellosis by enzyme-linked immunosorbent assay (ELISA) using an excretory-secretory antigen. *Vet. Parasitol.* 13, 349-361

Gamble HR, Rapic D, Marinculic A, Murrell KD. 1988. Evaluation of excretory-secretory antigens for the serodiagnosis of swine trichinellosis. *Vet. Parasitol.* 30, 131-137

OIE/World Organisation for Animal Health, 2018. Trichinellosis (infection with *Trichinella* spp.), Chapter 3.1.20. Manual of Diagnostic Tests and Vaccines for Terrestrial Animals. World Organization for Animal Health [www.oie.int/eng/normes/mmanual/A\\_00013.htm](http://www.oie.int/eng/normes/mmanual/A_00013.htm).

#### 3 Definitions

**OD**, optical density

**E/S Ag**, Excretory/Secretory antigens

#### 4 Devices/Instruments

Incubator  $37 \pm 2^\circ\text{C}$  with 4-5% CO<sub>2</sub>

Spectrophotometer UV/VIS

Freezer  $\leq -15^\circ\text{C}$

Laminar flow hood

Refrigerator,  $1^\circ\text{C} \div 8^\circ\text{C}$

Freezer  $\leq -50^\circ\text{C}$

Adjustable micropipettes (up to 2  $\mu\text{L}$ , 20  $\mu\text{L}$ , 200  $\mu\text{L}$ , 1000  $\mu\text{L}$ )

Pipettes (1, 5, 10, 25 mL)

Inverted microscope

Magnetic stirrer

Refrigerate centrifuge

96 wells plates

#### 5. Reagents and chemicals

##### 5.1 Phosphate Buffered Saline (PBS), pH $7.3 \pm 0.2$

KH <sub>2</sub> PO <sub>4</sub>	0.34 g
Na <sub>2</sub> HPO <sub>4</sub>	1.21 g
NaCl	8.0 g
Analytical grade water	up to 1000 mL

Dissolve the compounds in 750 mL of analytical grade water under magnetic stirring. Control the pH (it must be  $7.2 \pm 0.2$ ) and then bring the solution to the final volume. Sterilize by filtration with a 0.22  $\mu\text{m}$  filter. If stored refrigerated, it remains stable for 6 month.

- 5.2 PBS, pH 7.3 ± 0.2 with antibiotics 5 X
- |  |        |
|--|--------|
| PBS  | 950 mL |
| Penicillin/Streptomycin or Antibiotic/Antimycotic solution | 50 mL  |
- Store refrigerated and use within 2 months.
- 5.3 RPMI 1640 with antibiotics 5 X
- |  |        |
|--|--------|
| RPMI 1640  | 475 mL |
| Penicillin/Streptomycin or Antibiotic/Antimycotic solution | 25 mL  |
- Store refrigerated and use within 2 months.
- 5.4 Complete medium
- |   |        |
|---|--------|
| RPMI 1640   | 480 mL |
| 1M HEPES  | 5 mL   |
| 200mM Glutamine                                   | 5 mL   |
| 100mM Na-pyruvate                                 | 5 mL   |
| Penicillin/Streptomycin or Antibiotic/Antimycotic | 5 mL   |
- Store refrigerated and use within 2 months.
- 5.5 Penicillin/Streptomycin or Antibiotic/Antimycotic solution (100x).
- 5.6 RPMI 1640 culture medium.
- 5.7 HEPES-1M Buffer solution.
- 5.8 L-Glutamine 0.2M solution.
- 5.9 Sterile 0.22 µm syringe filters.
- 5.10 Device for concentrating by ultrafiltration with cellulose membrane filter, 5000 MWCO cut off (e.g. Amicon Ultra 15).
- 5.11 Coomassie Protein Assay Reagent (e.g. Pierce®, 1856209).
- 5.12 Slide-A-Lyzer Dialysis Cassettes, 3.5K MWCO, 3-12 ml Capacity (e.g. Pierce®, 66110).
- 5.13 Cocktail of protease inhibitors.
- 5.14 *Trichinella spiralis* muscle larvae suspension (MLS).
- 5.15 Sterile conical tubes 15, 50 mL.
- 5.16 Tissue culture flasks T-75.

## 6 Procedure

- Place the solutions 5.1, 5.2 and 5.3 into the 37°C incubator for at least 1 h.
- Count MSL under the inverted microscope.
- Wash 5X10<sup>5</sup> MSL three times (20 min each wash) by sedimentation in a sterile 50 ml conical tube with 45 mL of warm sterile PBS/Penicillin/Streptomycin solution (5.2). At each change of the washing solution, gently shake the worms to dislodge adherent bacteria. After 20 min of sedimentation, remove the washing solution by aspirating with a vacuum pump.
- Place the tube containing the worms under a laminar flow hood and wash them 5 additional times by sedimentation in a sterile 50 mL conical tube with 45 mL of warm RPMI 1640/Penicillin/Streptomycin solution (5.2).
- Resuspend the worms in warm maintenance media (5.4) at a concentration of 5,000 larvae/mL and place them into 25 mL tissue culture flasks.
- Incubate the flasks in 5% CO<sub>2</sub> at 37°C for 16-18 h.
- Check MSL viability and absence of bacterial contamination by microscopy.

- h) Place the cultures under a laminar flow hood and remove worms from the media by sedimentation in 50 mL conical tubes.
- i) Filter the media through a 0.2  $\mu\text{m}$  filter. Discard the MSL.
- j) Keep the E/S Ag at +4°C until concentration, if stored more than 24 hours, the E/S Ag should be frozen at -15°C.
- k) Fill in the concentration tube with 15 mL of E/S Ag (point i).
- l) Centrifuge 30 minutes at 3,000 g in a refrigerated centrifuge.
- m) Retrieve the concentrated E/S Ag in a 50 mL tube and store refrigerated until point “p”.
- n) Repeat point “k” to “m” until the E/S Ag has been properly concentrated.
- o) If the E/S Ag has been concentrated 100 times or more, proceed to point “p”, otherwise centrifuge the tube again.
- p) Determine the protein concentration by the Bradford method (5.11).
- q) Check the optical density by the spectrophotometer at 280nm/260nm. The ratio should be  $\geq 1$ .
- r) Dialyze the concentrated E/S Ag versus PBS at +4°C for at least 4 hours.
- s) Add to the E/S Ag 1 $\mu\text{L}$ /mg of the cocktail of protease inhibitors (5.13).
- t) Aliquot the antigen in 1 mL vials at a total protein concentration of 0.5 mg/vial and store at  $\leq -15^\circ\text{C}$ .
- u) The antigen can be lyophilized and stored refrigerated for 5 years..

## 7. Quality Control

The batch is considered suitable for serology if all the following criteria are fulfilled:

- Lack of microbial and fungal contamination in the culture flasks at the end of the incubation period, verified by direct microscopic observation at 400x magnifications, as specified in point 6 ‘g’;
- MSL viability, controlled as described in point 6 ‘g’;
- Optical density ratio of 280nm/260nm of the final antigen solution  $> 1$ , as specified in point 6 ‘q’.

## 8. Safety

This method has to be carried out only by authorized personnel. The operator should wear individual protection devices during the test performance. For the general safety measures, refer to the guidelines of CDC.