



IZS of Sardinia  
ITALY

## Thirteenth Workshop of National Reference Laboratories for Parasites

24-25 May, 2018



 Reference Laboratory  
for Echinococcosis

# VALIDATION OF A METHOD FOR THE DIAGNOSIS OF *E. GRANULOSUS* AND *E. MULTILOCULARIS* IN HUMANS BY MEANS OF A COMMERCIAL ELISA

Santucci C, Mura A, Peruzzu A, Masu G, Piseddu T, Bonelli P, Masala G.

[cinzia.santucci@izs-sardegna.it](mailto:cinzia.santucci@izs-sardegna.it)

Evaluation of several parameters determined to assess the suitability for the validation of a diagnostic method for the detection of IgG antibodies anti *E. granulosus* and *E. multilocularis* useful for a screening of human sera. Suitability of the parameters estimated carry to accreditation of the method.

Trials were performed according with the OIE guidelines and the quality management system of IZSs based on ACCREDIA (Italian Accreditation Body) procedure (Regulation UNI CEI EN ISO/IEC 17025).

## MATERIALS AND METHODS

For this purpose we employed the commercial ELISA kit “*Echinococcus IgG*” (DRG, Germany).

A total of 180 serum samples were tested:


80 belonged to people infected with *E. granulosus* (n=70) or *E. multilocularis* (n=10), 80 to *Echinococcus* free

healthy donors and 20 to people infected with *Trichinella britovi*, *Opisthorchis felineus*, *Tenia solium*, *Filarioidea*. (Neg samples from our Lab – EC Pos from Bio-Bank – sera pos to other parasites were provided by ISS - thank to Adriano Casulli and Maria Angeles Morales Gomez (EURLP)).

Diagnostic sensitivity (**DSe**), specificity (**DSp**), positive (**PPV**) and negative predictive values (**PNV**) were determined by a 2x2 contingency table by testing a statistical representative number of sera.

Moreover, analytical specificity (**Asp**) was assessed testing 30 sera divided in 5 groups, each of these infected by other parasites, analysed besides to *E. granulosus* and a negative control group.

The **repeatability** and **accuracy**, were evaluated by 3 operators that tested in triplicate a blind set of 20 samples, the agreement was calculated by Cohen-K, as well as, the **reproducibility**, evaluated by using 2 different sets of instruments, by 2 operators the analysed of a blind set of 27 samples.



# RESULTS



## DIAGNOSTIC SPECIFICITY AND SENSITIVITY (DSp – DSe)

		Clinical signs		Total	
		Presence of cysts	Absence of cysts		
Test Results	Pos	80	10	90	PPV 88.9
		(VP)	(FP)	(VP+FP)	VP/(VP+FP)
	Neg	0	90	90	PNV 100
		(FN)	(VN)	(FN+VN)	VN/(VN+FN)
Total		80	100	180	
		(VP+FN)	(FP+VN)	(VP+FP+FN+VN)	
		DSe 100	DSp 90		
		VP/(VP+FN)	VN/(VN+FP)		

## ANALYTICAL SPECIFICITY (Asp)

PARASITE SPECIES	ELISA RESULTS	
	positive	negative
<i>Echinococcus granulosus</i>	5/5	0/5
<i>Trichinella britovi</i>	1/5	4/5
<i>Opisthorchis felineus</i>	2/5	3/5
<i>Filariasi</i>	5/5	0/5
<i>Tenia solium</i>	2/5	3/5

## ANALYTICAL SPECIFICITY (Asp)

		Clinical signs		Total	
		Presence of cysts	Absence of cysts		
Test Results	Pos	5	10	15	PPV 33,3
		(VP)	(FP)	(VP+FP)	VP/(VP+FP)
	Neg	0	15	15	PNV 100
		(FN)	(VN)	(FN+VN)	VN/(VN+FN)
Total		5	25	30	
		(VP+FN)	(FP+VN)	(VP+FP+FN+VN)	
		ASe 100	Asp 60		
		VP/(VP+FN)	VN/(VN+FP)		

- **Repeatability** and **Accuracy** (operator performance)
  - **Reproducibility** (different instruments)
- excellent (k=1.00) agreement of the results

# CONCLUSIONS

According to the results of the parameters evaluated for validation, the ELISA method “*Echinococcus IgG*” showed a good performance, demonstrating that it is suitable for the accreditation .